

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121821\
 Data File : VW021480.D
 Acq On : 18 Dec 2021 16:42
 Operator : SY/VA
 Sample : M5154-12
 Misc : 5.44g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_W
ClientSampleId :
 BGL92

Quant Time: Dec 18 23:12:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120921SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Dec 18 00:55:20 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	166821	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.634	117	163372	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	96327	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	45961	15.128	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	60.520%	
7) Chloroethane-d5	2.879	69	28997	14.437	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	57.760%	
11) 1,1-Dichloroethene-d2	4.013	63	56316	13.903	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery	=	55.600%	
21) 2-Butanone-d5	7.080	46	21890	51.112	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	102.220%	
24) Chloroform-d	7.653	84	85216	18.330	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	=	73.320%	
26) 1,2-Dichloroethane-d4	8.305	65	51269	20.732	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery	=	82.920%	
32) Benzene-d6	8.274	84	170324	19.020	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	=	76.080%	
36) 1,2-Dichloropropane-d6	9.274	67	48151	20.049	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	=	80.200%	
41) Toluene-d8	10.323	98	169187	18.491	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	=	73.960%	
43) trans-1,3-Dichloroprop...	10.579	79	21837	19.747	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	=	79.000%	
47) 2-Hexanone-d5	10.920	63	18940	53.332	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	106.660%	
56) 1,1,2,2-Tetrachloroeth...	12.688	84	42646	20.656	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	=	82.640%	
66) 1,2-Dichlorobenzene-d4	13.853	152	68439	18.332	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	=	73.320%#	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121821\
 Data File : VW021480.D
 Acq On : 18 Dec 2021 16:42
 Operator : SY/VA
 Sample : M5154-12
 Misc : 5.44g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BGL92

Quant Time: Dec 18 23:12:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120921SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Dec 18 00:55:20 2021
 Response via : Initial Calibration

