

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121823\
 Data File : VW027735.D
 Acq On : 18 Dec 2023 17:07
 Operator : SY/MD
 Sample : 05794-11
 Misc : 5.58g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BH3Q6

Quant Time: Dec 18 22:14:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121523SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sun Dec 17 01:25:09 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	610019	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	478083	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	174869	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	189204	21.900	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	87.600%
7) Chloroethane-d5	2.899	69	156066	23.202	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	92.800%
11) 1,1-Dichloroethene-d2	4.021	65	76241	22.035	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	88.120%
21) 2-Butanone-d5	7.087	46	56228	38.221	ug/L	-0.01
Spiked Amount	50.000	Range	20 - 135	Recovery	=	76.440%
24) Chloroform-d	7.648	84	339446	23.385	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	93.520%
26) 1,2-Dichloroethane-d4	8.307	65	166535	24.435	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	97.720%
32) Benzene-d6	8.270	84	731792	26.191	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	104.760%
36) 1,2-Dichloropropane-d6	9.270	67	210804	26.631	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	106.520%
41) Toluene-d8	10.319	98	640098	24.692	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	98.760%
43) trans-1,3-Dichloroprop...	10.575	79	68023	22.195	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	88.800%
47) 2-Hexanone-d5	10.922	63	49539	42.944	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	85.880%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	142857	23.864	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	95.440%
66) 1,2-Dichlorobenzene-d4	13.848	152	144989	23.163	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	92.640%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121823\
Data File : VW027735.D
Acq On : 18 Dec 2023 17:07
Operator : SY/MD
Sample : 05794-11
Misc : 5.58g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BH3Q6

Quant Time: Dec 18 22:14:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121523SMA.M
Quant Title : SFAM01.0
QLast Update : Sun Dec 17 01:25:09 2023
Response via : Initial Calibration

