

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121824\
 Data File : VW031339.D
 Acq On : 18 Dec 2024 14:51
 Operator : SY/MD
 Sample : P5265-07
 Misc : 5.18g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Dec 19 00:24:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121324SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Dec 19 00:20:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	1257372	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	744241	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	171371	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.357	65	245650	12.985	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	51.920%
7) Chloroethane-d5	2.893	69	214242	15.309	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	61.240%
11) 1,1-Dichloroethene-d2	4.015	65	102290	11.398	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	45.600%
21) 2-Butanone-d5	7.100	46	19164	5.379	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	10.760%#
24) Chloroform-d	7.642	84	474114	13.762	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	55.040%
26) 1,2-Dichloroethane-d4	8.301	65	260401	15.051	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	60.200%#
32) Benzene-d6	8.276	84	724000	15.178	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	60.720%
36) 1,2-Dichloropropane-d6	9.270	67	272642	19.611	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	78.440%
41) Toluene-d8	10.318	98	492144	11.090	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	44.360%
43) trans-1,3-Dichloroprop...	10.575	79	33037	5.154	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	20.600%#
47) 2-Hexanone-d5	10.928	63	11773	6.297	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	12.600%#
56) 1,1,2,2-Tetrachloroeth...	12.690	84	144671	16.574	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	66.280%
66) 1,2-Dichlorobenzene-d4	13.848	152	70067	11.013	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	44.040%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121824\
Data File : VW031339.D
Acq On : 18 Dec 2024 14:51
Operator : SY/MD
Sample : P5265-07
Misc : 5.18g/10mL/MSVOA_W/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :

Quant Time: Dec 19 00:24:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121324SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Dec 19 00:20:44 2024
Response via : Initial Calibration

