

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121824\
 Data File : VW031353.D
 Acq On : 18 Dec 2024 21:16
 Operator : SY/MD
 Sample : P5302-02
 Misc : 7.36g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 E2AR1

Quant Time: Dec 19 04:11:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121324SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Dec 19 04:08:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	915377	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	707725	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	241394	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	263749	19.150	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	76.600%	
7) Chloroethane-d5	2.899	69	199087	19.541	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	78.160%	
11) 1,1-Dichloroethene-d2	4.015	65	117871	18.042	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery	=	72.160%	
21) 2-Butanone-d5	7.093	46	124964	48.178	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	96.360%	
24) Chloroform-d	7.648	84	453559	18.084	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	=	72.320%	
26) 1,2-Dichloroethane-d4	8.301	65	268891	21.348	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery	=	85.400%	
32) Benzene-d6	8.270	84	929101	20.482	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	=	81.920%	
36) 1,2-Dichloropropane-d6	9.270	67	268926	20.342	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	=	81.360%	
41) Toluene-d8	10.318	98	802938	19.027	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	=	76.120%	
43) trans-1,3-Dichloroprop...	10.575	79	113716	18.655	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	=	74.600%	
47) 2-Hexanone-d5	10.922	63	98427	55.362	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	110.720%	
56) 1,1,2,2-Tetrachloroeth...	12.690	84	191746	23.100	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	=	92.400%	
66) 1,2-Dichlorobenzene-d4	13.848	152	169033	18.862	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	=	75.440%	
Target Compounds					Qvalue	
46) Tetrachloroethene	10.861	164	129405	14.308	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121824\
Data File : VW031353.D
Acq On : 18 Dec 2024 21:16
Operator : SY/MD
Sample : P5302-02
Misc : 7.36g/10mL/MSVOA_W/SOIL/A
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
E2AR1

Quant Time: Dec 19 04:11:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121324SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Dec 19 04:08:34 2024
Response via : Initial Calibration

