

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121824\
 Data File : VW031358.D
 Acq On : 18 Dec 2024 23:17
 Operator : SY/MD
 Sample : P5302-07
 Misc : 4.90g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 E2AS0

Quant Time: Dec 19 04:12:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121324SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Dec 19 04:08:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	958695	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	609277	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.550	152	162691	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.357	65	250675	17.378	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	69.520%
7) Chloroethane-d5	2.893	69	200143	18.757	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	75.040%
11) 1,1-Dichloroethene-d2	4.015	65	105045	15.352	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	61.400%
21) 2-Butanone-d5	7.100	46	45091	16.599	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	33.200%
24) Chloroform-d	7.648	84	424368	16.156	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	64.640%
26) 1,2-Dichloroethane-d4	8.307	65	246213	18.664	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	74.640%
32) Benzene-d6	8.270	84	767790	19.661	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	78.640%
36) 1,2-Dichloropropane-d6	9.270	67	253715	22.292	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	89.160%
41) Toluene-d8	10.319	98	565538	15.567	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	62.280%
43) trans-1,3-Dichloroprop...	10.575	79	61396	11.699	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	46.800%
47) 2-Hexanone-d5	10.922	63	27743	18.126	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	36.260%
56) 1,1,2,2-Tetrachloroeth...	12.684	84	147406	20.628	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	82.520%
66) 1,2-Dichlorobenzene-d4	13.849	152	94499	15.646	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	62.600%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121824\
Data File : VW031358.D
Acq On : 18 Dec 2024 23:17
Operator : SY/MD
Sample : P5302-07
Misc : 4.90g/10mL/MSVOA_W/SOIL/A
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
E2AS0

Quant Time: Dec 19 04:12:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121324SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Dec 19 04:08:34 2024
Response via : Initial Calibration

