

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121824\
 Data File : VW031331.D
 Acq On : 18 Dec 2024 11:38
 Operator : SY/MD
 Sample : P5155-08
 Misc : 5.04g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK001

Manual Integrations
 APPROVED

Reviewed By :Romaben Patel 12/19/2024
 Supervised By :Mahesh Dadoda 12/20/2024

Quant Time: Dec 19 00:21:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121324SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Dec 19 00:20:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	653672	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	566228	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	278361	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.357	65	220387	22.408	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	89.640%
7) Chloroethane-d5	2.899	69	171055	23.511	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	94.040%
11) 1,1-Dichloroethene-d2	4.015	65	101137	21.678	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	86.720%
21) 2-Butanone-d5	7.094	46	94833	51.200	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	102.400%
24) Chloroform-d	7.642	84	416799	23.272	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	93.080%
26) 1,2-Dichloroethane-d4	8.301	65	229547m	25.520	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	102.080%
32) Benzene-d6	8.270	84	850130	23.425	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	93.680%
36) 1,2-Dichloropropane-d6	9.270	67	246018	23.259	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	93.040%
41) Toluene-d8	10.319	98	772411	22.878	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	91.520%
43) trans-1,3-Dichloroprop...	10.575	79	107631	22.069	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	88.280%
47) 2-Hexanone-d5	10.922	63	73714	51.822	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	103.640%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	176740	26.613	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	106.440%
66) 1,2-Dichlorobenzene-d4	13.848	152	250520	24.243	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	96.960%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121824\
Data File : VW031331.D
Acq On : 18 Dec 2024 11:38
Operator : SY/MD
Sample : P5155-08
Misc : 5.04g/10mL/MSVOA_W/SOIL/A
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK001

Quant Time: Dec 19 00:21:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121324SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Dec 19 00:20:44 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Romaben Patel 12/19/2024
Supervised By :Mahesh Dadoda 12/20/2024

