

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_W\Data\VW121824\  
 Data File : VW031339.D  
 Acq On : 18 Dec 2024 14:51  
 Operator : SY/MD  
 Sample : P5265-07  
 Misc : 5.18g/10mL/MSVOA\_W/SOIL/A  
 ALS Vial : 13 Sample Multiplier: 1

Instrument :  
 MSVOA\_W  
 ClientSampleId :  
 E28Z2

Quant Time: Dec 19 00:24:01 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_W\Method\SFAMWLM121324SMA.M  
 Quant Title : SFAM01.0  
 QLast Update : Thu Dec 19 00:20:44 2024  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-------------------------------|--------|-------|----------|----------|-------|----------|
| -----                         |        |       |          |          |       |          |
| Internal Standards            |        |       |          |          |       |          |
| 1) 1,4-Difluorobenzene        | 8.837  | 114   | 1257372  | 25.000   | ug/L  | 0.00     |
| 28) Chlorobenzene-d5          | 11.629 | 117   | 744241   | 25.000   | ug/L  | 0.00     |
| 58) 1,4-Dichlorobenzene-d4    | 13.556 | 152   | 171371   | 25.000   | ug/L  | 0.00     |
|                               |        |       |          |          |       |          |
| System Monitoring Compounds   |        |       |          |          |       |          |
| 4) Vinyl Chloride-d3          | 2.357  | 65    | 245650   | 12.985   | ug/L  | 0.00     |
| Spiked Amount                 | 25.000 | Range | 30 - 150 | Recovery | =     | 51.920%  |
| 7) Chloroethane-d5            | 2.893  | 69    | 214242   | 15.309   | ug/L  | 0.00     |
| Spiked Amount                 | 25.000 | Range | 30 - 150 | Recovery | =     | 61.240%  |
| 11) 1,1-Dichloroethene-d2     | 4.015  | 65    | 102290   | 11.398   | ug/L  | 0.00     |
| Spiked Amount                 | 25.000 | Range | 45 - 110 | Recovery | =     | 45.600%  |
| 21) 2-Butanone-d5             | 7.100  | 46    | 19164    | 5.379    | ug/L  | 0.00     |
| Spiked Amount                 | 50.000 | Range | 20 - 135 | Recovery | =     | 10.760%# |
| 24) Chloroform-d              | 7.642  | 84    | 474114   | 13.762   | ug/L  | 0.00     |
| Spiked Amount                 | 25.000 | Range | 40 - 150 | Recovery | =     | 55.040%  |
| 26) 1,2-Dichloroethane-d4     | 8.301  | 65    | 260401   | 15.051   | ug/L  | 0.00     |
| Spiked Amount                 | 25.000 | Range | 70 - 130 | Recovery | =     | 60.200%# |
| 32) Benzene-d6                | 8.276  | 84    | 724000   | 15.178   | ug/L  | 0.00     |
| Spiked Amount                 | 25.000 | Range | 20 - 135 | Recovery | =     | 60.720%  |
| 36) 1,2-Dichloropropane-d6    | 9.270  | 67    | 272642   | 19.611   | ug/L  | 0.00     |
| Spiked Amount                 | 25.000 | Range | 70 - 120 | Recovery | =     | 78.440%  |
| 41) Toluene-d8                | 10.318 | 98    | 492144   | 11.090   | ug/L  | 0.00     |
| Spiked Amount                 | 25.000 | Range | 30 - 130 | Recovery | =     | 44.360%  |
| 43) trans-1,3-Dichloroprop... | 10.575 | 79    | 33037    | 5.154    | ug/L  | 0.00     |
| Spiked Amount                 | 25.000 | Range | 30 - 135 | Recovery | =     | 20.600%# |
| 47) 2-Hexanone-d5             | 10.928 | 63    | 11773    | 6.297    | ug/L  | 0.00     |
| Spiked Amount                 | 50.000 | Range | 20 - 135 | Recovery | =     | 12.600%# |
| 56) 1,1,2,2-Tetrachloroeth... | 12.690 | 84    | 144671   | 16.574   | ug/L  | 0.00     |
| Spiked Amount                 | 25.000 | Range | 45 - 120 | Recovery | =     | 66.280%  |
| 66) 1,2-Dichlorobenzene-d4    | 13.848 | 152   | 70067    | 11.013   | ug/L  | 0.00     |
| Spiked Amount                 | 25.000 | Range | 75 - 120 | Recovery | =     | 44.040%# |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_W\Data\VW121824\  
Data File : VW031339.D  
Acq On : 18 Dec 2024 14:51  
Operator : SY/MD  
Sample : P5265-07  
Misc : 5.18g/10mL/MSVOA\_W/SOIL/A  
ALS Vial : 13 Sample Multiplier: 1

Instrument :  
MSVOA\_W  
ClientSampleId :  
E28Z2

Quant Time: Dec 19 00:24:01 2024  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_W\Method\SFAMWLM121324SMA.M  
Quant Title : SFAM01.0  
QLast Update : Thu Dec 19 00:20:44 2024  
Response via : Initial Calibration

