

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121824\
 Data File : VW031341.D
 Acq On : 18 Dec 2024 15:39
 Operator : SY/MD
 Sample : P5265-09
 Misc : 2.43g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 E28Z4

Quant Time: Dec 19 00:24:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121324SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Dec 19 00:20:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	1150880	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	621733	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	136150	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	230996	13.340	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	53.360%
7) Chloroethane-d5	2.893	69	197823	15.444	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	61.760%
11) 1,1-Dichloroethene-d2	4.015	65	95313	11.604	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	46.400%
21) 2-Butanone-d5	7.112	46	10201	3.128	ug/L	0.02
Spiked Amount	50.000	Range	20 - 135	Recovery	=	6.260%#
24) Chloroform-d	7.648	84	416462	13.207	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	52.840%
26) 1,2-Dichloroethane-d4	8.301	65	234994	14.839	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	59.360%#
32) Benzene-d6	8.270	84	659814	16.558	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	66.240%
36) 1,2-Dichloropropane-d6	9.270	67	240564	20.713	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	82.840%
41) Toluene-d8	10.318	98	459801	12.403	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	49.600%
43) trans-1,3-Dichloroprop...	10.575	79	25338	4.732	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	18.920%#
47) 2-Hexanone-d5	10.928	63	5888	3.770	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	7.540%#
56) 1,1,2,2-Tetrachloroeth...	12.690	84	123357	16.917	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	67.680%
66) 1,2-Dichlorobenzene-d4	13.854	152	55756	11.031	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	44.120%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121824\
Data File : VW031341.D
Acq On : 18 Dec 2024 15:39
Operator : SY/MD
Sample : P5265-09
Misc : 2.43g/10mL/MSVOA_W/SOIL/A
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
E28Z4

Quant Time: Dec 19 00:24:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121324SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Dec 19 00:20:44 2024
Response via : Initial Calibration

