

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121824\
 Data File : VW031351.D
 Acq On : 18 Dec 2024 20:04
 Operator : SY/MD
 Sample : VW1218SBL02
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK410

Quant Time: Dec 19 04:10:22 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121324SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Dec 19 04:08:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	620350	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	555559	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	274571	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	174679	18.714	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	74.840%	
7) Chloroethane-d5	2.899	69	140193	20.304	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	81.200%	
11) 1,1-Dichloroethene-d2	4.015	65	79952	18.058	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery	=	72.240%	
21) 2-Butanone-d5	7.093	46	89890	51.138	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	102.280%	
24) Chloroform-d	7.648	84	351945	20.706	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	=	82.840%	
26) 1,2-Dichloroethane-d4	8.307	65	202714	23.748	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery	=	95.000%	
32) Benzene-d6	8.270	84	720895	20.245	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	=	81.000%	
36) 1,2-Dichloropropane-d6	9.270	67	214684	20.687	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	=	82.760%	
41) Toluene-d8	10.319	98	660331	19.934	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	=	79.720%	
43) trans-1,3-Dichloroprop...	10.575	79	92296	19.288	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	=	77.160%	
47) 2-Hexanone-d5	10.922	63	71936	51.544	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	103.080%	
56) 1,1,2,2-Tetrachloroeth...	12.690	84	165880	25.458	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	=	101.840%	
66) 1,2-Dichlorobenzene-d4	13.848	152	226038	22.175	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	=	88.720%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121824\
Data File : VW031351.D
Acq On : 18 Dec 2024 20:04
Operator : SY/MD
Sample : VW1218SBL02
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK410

Quant Time: Dec 19 04:10:22 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121324SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Dec 19 04:08:34 2024
Response via : Initial Calibration

