

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121924\
 Data File : VW031401.D
 Acq On : 19 Dec 2024 22:56
 Operator : SY/MD
 Sample : P5333-03
 Misc : 6.89g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 E2906

Quant Time: Dec 20 04:53:10 2024
 Quant Title :
 QLast Update : Mon Dec 23 01:56:19 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	175787	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	96407	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	22096	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.357	65	62229	23.528	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	94.120%
7) Chloroethane-d5	2.893	69	46043	23.533	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	94.120%
11) 1,1-Dichloroethene-d2	4.015	65	25540	20.357	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	81.440%
21) 2-Butanone-d5	7.088	46	9099	18.267	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	36.540%
24) Chloroform-d	7.648	84	82998	17.232	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	68.920%
26) 1,2-Dichloroethane-d4	8.307	65	44319	18.322	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	73.280%
32) Benzene-d6	8.276	84	159949	25.885	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	103.560%
36) 1,2-Dichloropropane-d6	9.270	67	48750	27.070	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	108.280%
41) Toluene-d8	10.319	98	111762	19.442	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	77.760%
43) trans-1,3-Dichloroprop...	10.575	79	10193	12.275	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	49.120%
47) 2-Hexanone-d5	10.928	63	7577	31.286	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	62.580%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	27406	24.238	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	96.960%
66) 1,2-Dichlorobenzene-d4	13.849	152	14617	17.819	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	71.280%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121924\
Data File : VW031401.D
Acq On : 19 Dec 2024 22:56
Operator : SY/MD
Sample : P5333-03
Misc : 6.89g/10mL/MSVOA_W/SOIL/A
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
E2906

Quant Time: Dec 20 04:53:10 2024
Quant Title :
QLast Update : Mon Dec 23 01:56:19 2024
Response via : Initial Calibration

