

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121924\
 Data File : VW031371.D
 Acq On : 19 Dec 2024 09:34
 Operator : SY/MD
 Sample : VW1219SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK411

Quant Time: Dec 20 00:27:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121324SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Dec 20 00:25:15 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	621012	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	546095	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.550	152	255038	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.357	65	204509	21.887	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	87.560%
7) Chloroethane-d5	2.899	69	164500	23.799	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	95.200%
11) 1,1-Dichloroethene-d2	4.015	65	95589	21.567	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	86.280%
21) 2-Butanone-d5	7.094	46	97661	55.499	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	111.000%
24) Chloroform-d	7.648	84	403597	23.720	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	94.880%
26) 1,2-Dichloroethane-d4	8.301	65	224653	26.290	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	105.160%
32) Benzene-d6	8.270	84	816779	23.335	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	93.360%
36) 1,2-Dichloropropane-d6	9.270	67	241369	23.661	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	94.640%
41) Toluene-d8	10.319	98	747214	22.947	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	91.800%
43) trans-1,3-Dichloroprop...	10.575	79	101807	21.644	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	86.560%
47) 2-Hexanone-d5	10.922	63	75758	55.223	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	110.440%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	178209	27.824	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	111.280%
66) 1,2-Dichlorobenzene-d4	13.849	152	244741	25.849	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	103.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121924\
Data File : VW031371.D
Acq On : 19 Dec 2024 09:34
Operator : SY/MD
Sample : VW1219SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK411

Quant Time: Dec 20 00:27:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121324SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Dec 20 00:25:15 2024
Response via : Initial Calibration

