

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121924\
 Data File : VW031373.D
 Acq On : 19 Dec 2024 10:32
 Operator : SY/MD
 Sample : P5265-01RE
 Misc : 6.81g/10mL/MSVOA_W/SOIL/B
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 E28Y6RE

Quant Time: Dec 20 02:52:55 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121324SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Dec 20 00:25:15 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	469753	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	173787	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	30291	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.357	65	203403	28.778	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	= 115.120%		
7) Chloroethane-d5	2.887	69	150392	28.765	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	= 115.040%		
11) 1,1-Dichloroethene-d2	4.015	65	60023	17.903	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery	= 71.600%		
21) 2-Butanone-d5	7.093	46	100331	75.376	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	= 150.760%#		
24) Chloroform-d	7.642	84	268113	20.831	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	= 83.320%		
26) 1,2-Dichloroethane-d4	8.301	65	155705	24.088	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery	= 96.360%		
32) Benzene-d6	8.270	84	495338	44.470	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	= 177.880%#		
36) 1,2-Dichloropropane-d6	9.270	67	167509	51.599	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	= 206.400%#		
41) Toluene-d8	10.319	98	256380	24.741	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	= 98.960%		
43) trans-1,3-Dichloroprop...	10.575	79	27880	18.626	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	= 74.520%		
47) 2-Hexanone-d5	10.922	63	59294	135.816	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	= 271.640%#		
56) 1,1,2,2-Tetrachloroeth...	12.690	84	117808	57.798	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	= 231.200%#		
66) 1,2-Dichlorobenzene-d4	13.848	152	26321	23.406	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	= 93.640%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121924\
Data File : VW031373.D
Acq On : 19 Dec 2024 10:32
Operator : SY/MD
Sample : P5265-01RE
Misc : 6.81g/10mL/MSVOA_W/SOIL/B
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
E28Y6RE

Quant Time: Dec 20 02:52:55 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121324SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Dec 20 00:25:15 2024
Response via : Initial Calibration

