

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121924\
 Data File : VW031375.D
 Acq On : 19 Dec 2024 11:20
 Operator : SY/MD
 Sample : P5265-03
 Misc : 4.38g/10mL/MSVOA_W/SOIL/B
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 E28Y8

Quant Time: Dec 20 00:28:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121324SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Dec 20 00:25:15 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	723431	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	619861	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	299369	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.357	65	190147	17.469	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	69.880%	
7) Chloroethane-d5	2.893	69	142276	17.670	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	70.680%	
11) 1,1-Dichloroethene-d2	4.015	65	89506	17.335	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery	=	69.360%	
21) 2-Butanone-d5	7.100	46	62873	30.671	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	61.340%	
24) Chloroform-d	7.642	84	324201	16.356	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	=	65.440%	
26) 1,2-Dichloroethane-d4	8.307	65	173298	17.409	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery	=	69.640%#	
32) Benzene-d6	8.270	84	689871	17.364	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	=	69.440%	
36) 1,2-Dichloropropane-d6	9.270	67	194482	16.796	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	=	67.200%#	
41) Toluene-d8	10.319	98	640962	17.342	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	=	69.360%	
43) trans-1,3-Dichloroprop...	10.575	79	77061	14.434	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	=	57.720%	
47) 2-Hexanone-d5	10.922	63	49743	31.945	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	63.880%	
56) 1,1,2,2-Tetrachloroeth...	12.690	84	125386	17.247	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	=	69.000%	
66) 1,2-Dichlorobenzene-d4	13.848	152	186970	16.823	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	=	67.280%#	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121924\
Data File : VW031375.D
Acq On : 19 Dec 2024 11:20
Operator : SY/MD
Sample : P5265-03
Misc : 4.38g/10mL/MSVOA_W/SOIL/B
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
E28Y8

Quant Time: Dec 20 00:28:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121324SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Dec 20 00:25:15 2024
Response via : Initial Calibration

