

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121924\
 Data File : VW031377.D
 Acq On : 19 Dec 2024 12:08
 Operator : SY/MD
 Sample : P5265-05
 Misc : 2.65g/10mL/MSVOA_W/SOIL/B
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 E28Z0

Quant Time: Dec 20 00:29:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121324SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Dec 20 00:25:15 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	687037	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	597389	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.550	152	288802	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.357	65	208509	20.171	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	80.680%	
7) Chloroethane-d5	2.893	69	157017	20.534	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	82.120%	
11) 1,1-Dichloroethene-d2	4.015	65	95664	19.509	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery	=	78.040%	
21) 2-Butanone-d5	7.100	46	72195	37.085	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	74.160%	
24) Chloroform-d	7.648	84	361064	19.181	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	=	76.720%	
26) 1,2-Dichloroethane-d4	8.307	65	192626	20.375	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery	=	81.520%	
32) Benzene-d6	8.270	84	761490	19.888	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	=	79.560%	
36) 1,2-Dichloropropane-d6	9.270	67	213193	19.105	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	=	76.400%	
41) Toluene-d8	10.319	98	696051	19.541	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	=	78.160%	
43) trans-1,3-Dichloroprop...	10.575	79	84349	16.393	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	=	65.560%	
47) 2-Hexanone-d5	10.922	63	57067	38.027	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	76.060%	
56) 1,1,2,2-Tetrachloroeth...	12.690	84	144231	20.585	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	=	82.360%	
66) 1,2-Dichlorobenzene-d4	13.848	152	219512	20.474	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	=	81.880%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121924\
Data File : VW031377.D
Acq On : 19 Dec 2024 12:08
Operator : SY/MD
Sample : P5265-05
Misc : 2.65g/10mL/MSVOA_W/SOIL/B
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
E28Z0

Quant Time: Dec 20 00:29:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121324SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Dec 20 00:25:15 2024
Response via : Initial Calibration

