

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121924\
 Data File : VW031398.D
 Acq On : 19 Dec 2024 21:44
 Operator : SY/MD
 Sample : P5302-10
 Misc : 5.22g/10mL/MSVOA_W/SOIL/B
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 E2AS1

Quant Time: Dec 20 04:37:16 2024
 Quant Title :
 QLast Update : Mon Dec 23 01:56:19 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	1035707	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	824016	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	306648	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	259468	16.650	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	66.600%
7) Chloroethane-d5	2.899	69	195559	16.965	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	67.840%
11) 1,1-Dichloroethene-d2	4.015	65	111672	15.107	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	60.440%
21) 2-Butanone-d5	7.093	46	91484	31.173	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	62.340%
24) Chloroform-d	7.648	84	407326	14.354	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	57.400%
26) 1,2-Dichloroethane-d4	8.307	65	228346	16.022	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	64.080%#
32) Benzene-d6	8.270	84	847029	16.038	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	64.160%
36) 1,2-Dichloropropane-d6	9.270	67	250093	16.248	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	65.000%#
41) Toluene-d8	10.319	98	723970	14.735	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	58.920%
43) trans-1,3-Dichloroprop...	10.575	79	96350	13.575	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	54.320%
47) 2-Hexanone-d5	10.922	63	72383	34.967	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	69.940%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	171876	17.784	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	71.120%
66) 1,2-Dichlorobenzene-d4	13.854	152	173531	15.243	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	60.960%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121924\
Data File : VW031398.D
Acq On : 19 Dec 2024 21:44
Operator : SY/MD
Sample : P5302-10
Misc : 5.22g/10mL/MSVOA_W/SOIL/B
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
E2AS1

Quant Time: Dec 20 04:37:16 2024
Quant Title :
QLast Update : Mon Dec 23 01:56:19 2024
Response via : Initial Calibration

