

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121924\
 Data File : VW031400.D
 Acq On : 19 Dec 2024 22:32
 Operator : SY/MD
 Sample : P5333-02
 Misc : 3.02g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 E2905

Manual Integrations
 APPROVED

Reviewed By :Romaben Patel 12/20/2024
 Supervised By :Mahesh Dadoda 12/23/2024

Quant Time: Dec 20 04:48:43 2024
 Quant Title :
 QLast Update : Mon Dec 23 01:56:19 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	1051935	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	806324	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	297364	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.357	65	260801	16.478	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	65.920%
7) Chloroethane-d5	2.893	69	204623	17.477	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	69.920%
11) 1,1-Dichloroethene-d2	4.015	65	110986	14.783	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	59.120%
21) 2-Butanone-d5	7.100	46	43778	14.687	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	29.380%
24) Chloroform-d	7.648	84	434474	15.074	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	60.280%
26) 1,2-Dichloroethane-d4	8.307	65	228260m	15.769	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	63.080%#
32) Benzene-d6	8.270	84	856317	16.569	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	66.280%
36) 1,2-Dichloropropane-d6	9.270	67	259005	17.196	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	68.800%#
41) Toluene-d8	10.319	98	740718	15.406	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	61.640%
43) trans-1,3-Dichloroprop...	10.575	79	66382	9.558	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	38.240%
47) 2-Hexanone-d5	10.922	63	32183	15.888	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	31.780%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	142760	15.096	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	60.400%
66) 1,2-Dichlorobenzene-d4	13.848	152	159417	14.441	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	57.760%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121924\
Data File : VW031400.D
Acq On : 19 Dec 2024 22:32
Operator : SY/MD
Sample : P5333-02
Misc : 3.02g/10mL/MSVOA_W/SOIL/A
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
E2905

Quant Time: Dec 20 04:48:43 2024
Quant Title :
QLast Update : Mon Dec 23 01:56:19 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Romaben Patel 12/20/2024
Supervised By :Mahesh Dadoda 12/23/2024

