

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121924\
 Data File : VW031403.D
 Acq On : 19 Dec 2024 23:44
 Operator : SY/MD
 Sample : P5333-05
 Misc : 6.47g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 E2908

Manual Integrations
 APPROVED

Reviewed By :Romaben Patel 12/20/2024
 Supervised By :Mahesh Dadoda 12/23/2024

Quant Time: Dec 20 04:09:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121324SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Dec 20 03:08:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	728993	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	382264	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	73972	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.357	65	191104	17.423	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	69.680%		
7) Chloroethane-d5	2.893	69	147021	18.120	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	72.480%		
11) 1,1-Dichloroethene-d2	4.015	65	76845	14.770	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	59.080%		
21) 2-Butanone-d5	7.106	46	41072	19.883	ug/L	0.01
Spiked Amount 50.000	Range 20 - 135		Recovery =	39.760%		
24) Chloroform-d	7.648	84	286247	14.331	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	57.320%		
26) 1,2-Dichloroethane-d4	8.307	65	165886m	16.537	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	66.160%#		
32) Benzene-d6	8.276	84	572629	23.372	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	93.480%		
36) 1,2-Dichloropropane-d6	9.270	67	170947	23.940	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	95.760%		
41) Toluene-d8	10.319	98	393530	17.265	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	69.080%		
43) trans-1,3-Dichloroprop...	10.581	79	40730	12.370	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	49.480%		
47) 2-Hexanone-d5	10.922	63	31592	32.898	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	65.800%		
56) 1,1,2,2-Tetrachloroeth...	12.690	84	95674	21.340	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	85.360%		
66) 1,2-Dichlorobenzene-d4	13.848	152	38147	13.891	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	55.560%#		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121924\
Data File : VW031403.D
Acq On : 19 Dec 2024 23:44
Operator : SY/MD
Sample : P5333-05
Misc : 6.47g/10mL/MSVOA_W/SOIL/A
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
E2908

Manual Integrations
APPROVED

Reviewed By :Romaben Patel 12/20/2024
Supervised By :Mahesh Dadoda 12/23/2024

Quant Time: Dec 20 04:09:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121324SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Dec 20 03:08:41 2024
Response via : Initial Calibration

