

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW122018\
 Data File : VW007687.D
 Acq On : 19 Dec 2018 05:50
 Operator : SY/AP
 Sample : J6386-10
 Misc : 6.12G/5ML/MSVOA_W/SOIL
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 P001-SS006-0612-01

Quant Time: Dec 19 06:45:05 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W121118S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 11 02:58:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	30346	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	59977	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	45539	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	14277	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	29512	87.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	175.32%	
35) Dibromofluoromethane	7.88	113	21879	60.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	120.52%	
50) Toluene-d8	10.32	98	56646	41.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.22%	
62) 4-Bromofluorobenzene	12.62	95	16692	33.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	66.70%	
Target Compounds						
16) Acetone	4.12	43	26446	449.721	ug/l	99
20) Methylene Chloride	4.91	84	4600	11.270	ug/l	98
25) 2-Butanone	7.17	43	1589	20.698	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_W\DATA\VW122018\
Data File : VW007687.D
Acq On : 19 Dec 2018 05:50
Operator : SY/AP
Sample : J6386-10
Misc : 6.12G/5ML/MSVOA_W/SOIL
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
P001-SS006-0612-01

Quant Time: Dec 19 06:45:05 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W121118S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 11 02:58:16 2018
Response via : Initial Calibration

