

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW122118\
 Data File : VW007741.D
 Acq On : 20 Dec 2018 11:33
 Operator : SY/AP
 Sample : J6386-09RE
 Misc : 3.32G/5ML/MSVOA W/SOIL
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 P001-SS006-0006-01RE

Quant Time: Dec 20 11:57:46 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W122118S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 20 06:30:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	50371	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	78210	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	45153	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	11358	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	30886	55.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.88%	
35) Dibromofluoromethane	7.88	113	25948	54.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.76%	
50) Toluene-d8	10.32	98	61441	34.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	68.32%	
62) 4-Bromofluorobenzene	12.62	95	13759	19.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	39.60%	
Target Compounds						
16) Acetone	4.12	43	19299	186.673	ug/l	91
25) 2-Butanone	7.18	43	1175	9.703	ug/l #	86
52) Toluene	10.40	92	1425	1.054	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW122118\
Data File : VW007741.D
Acq On : 20 Dec 2018 11:33
Operator : SY/AP
Sample : J6386-09RE
Misc : 3.32G/5ML/MSVOA W/SOIL
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
P001-SS006-0006-01RE

Quant Time: Dec 20 11:57:46 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W122118S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 20 06:30:23 2018
Response via : Initial Calibration

