

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW122118\
 Data File : VW007721.D
 Acq On : 20 Dec 2018 02:30
 Operator : SY/AP
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 20 07:44:11 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W122118S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 20 06:30:23 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	102	0.00
2 T	Dichlorodifluoromethane	0.444	0.383	13.7	88	0.00
3 P	Chloromethane	0.273	0.303	-11.0	117	0.00
4 C	Vinyl Chloride	0.347	0.373	-7.5#	115	0.00
5 T	Bromomethane	0.242	0.269	-11.2	123	0.00
6 T	Chloroethane	0.186	0.213	-14.5	123	0.00
7 T	Trichlorofluoromethane	0.362	0.328	9.4	96	0.00
8 T	Diethyl Ether	0.222	0.241	-8.6	111	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.533	0.476	10.7	95	0.00
10 T	Methyl Iodide	0.847	0.849	-0.2	106	0.00
11 T	Tert butyl alcohol	0.037	0.037	0.0	112	0.00
12 CM	1,1-Dichloroethene	0.493	0.468	5.1#	101	0.00
13 T	Acrolein	0.030	0.027	10.0	102	0.00
14 T	Allyl chloride	0.647	0.670	-3.6	111	0.00
15 T	Acrylonitrile	0.095	0.094	1.1	104	0.00
16 T	Acetone	0.103	0.086	16.5	87	0.00
17 T	Carbon Disulfide	1.527	1.466	4.0	103	0.00
18 T	Methyl Acetate	0.244	0.253	-3.7	105	0.00
19 T	Methyl tert-butyl Ether	0.708	0.756	-6.8	110	0.00
20 T	Methylene Chloride	0.619	0.523	15.5	103	0.00
21 T	trans-1,2-Dichloroethene	0.543	0.527	2.9	106	0.00
22 T	Diisopropyl ether	1.211	1.329	-9.7	114	0.00
23 T	Vinyl Acetate	0.720	0.742	-3.1	107	0.00
24 P	1,1-Dichloroethane	0.891	0.899	-0.9	109	0.00
25 T	2-Butanone	0.120	0.112	6.7	99	0.00
26 T	2,2-Dichloropropane	0.652	0.593	9.0	100	0.00
27 T	cis-1,2-Dichloroethene	0.578	0.587	-1.6	108	0.00
28 T	Bromochloromethane	0.304	0.362	-19.1	126	0.00
29 T	Tetrahydrofuran	0.072	0.074	-2.8	107	0.00
30 C	Chloroform	1.008	1.006	0.2#	106	0.00
31 T	Cyclohexane	0.816	0.743	8.9	104	0.00
32 T	1,1,1-Trichloroethane	0.918	0.886	3.5	103	0.00
33 S	1,2-Dichloroethane-d4	0.548	0.577	-5.3	111	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	108	0.00
35 S	Dibromofluoromethane	0.305	0.324	-6.2	117	0.00
36 T	1,1-Dichloropropene	0.483	0.452	6.4	104	0.00
37 T	Ethyl Acetate	0.167	0.160	4.2	105	0.00
38 T	Carbon Tetrachloride	0.568	0.499	12.1	97	0.00
39 T	Methylcyclohexane	0.600	0.542	9.7	100	0.00
40 TM	Benzene	1.288	1.247	3.2	108	0.00
41 T	Methacrylonitrile	0.095	0.095	0.0	107	0.00
42 TM	1,2-Dichloroethane	0.432	0.400	7.4	102	0.00
43 T	Isopropyl Acetate	0.334	0.324	3.0	105	0.00
44 TM	Trichloroethene	0.389	0.366	5.9	106	0.00
45 C	1,2-Dichloropropane	0.286	0.288	-0.7#	109	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW122118\
 Data File : VW007721.D
 Acq On : 20 Dec 2018 02:30
 Operator : SY/AP
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 20 07:44:11 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W122118S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 20 06:30:23 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.176	0.166	5.7	104	0.00
47 T	Bromodichloromethane	0.462	0.448	3.0	106	0.00
48 T	Methyl methacrylate	0.167	0.165	1.2	108	0.00
49 T	1,4-Dioxane	0.003	0.003	0.0	106	0.00
50 S	Toluene-d8	1.150	1.244	-8.2	118	0.00
51 T	4-Methyl-2-Pentanone	0.158	0.152	3.8	105	0.00
52 CM	Toluene	0.864	0.842	2.5#	108	0.00
53 T	t-1,3-Dichloropropene	0.460	0.439	4.6	106	0.00
54 T	cis-1,3-Dichloropropene	0.508	0.495	2.6	108	0.00
55 T	1,1,2-Trichloroethane	0.242	0.234	3.3	109	0.00
56 T	Ethyl methacrylate	0.287	0.286	0.3	105	0.00
57 T	1,3-Dichloropropane	0.403	0.395	2.0	109	0.00
58 T	2-Chloroethyl Vinyl ether	0.131	0.146	-11.5	118	0.00
59 T	2-Hexanone	0.111	0.104	6.3	100	0.00
60 T	Dibromochloromethane	0.321	0.304	5.3	103	0.00
61 T	1,2-Dibromoethane	0.246	0.231	6.1	105	0.00
62 S	4-Bromofluorobenzene	0.444	0.475	-7.0	117	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	108	0.00
64 T	Tetrachloroethene	0.354	0.350	1.1	110	0.00
65 PM	Chlorobenzene	1.082	1.035	4.3	108	0.00
66 T	1,1,1,2-Tetrachloroethane	0.387	0.373	3.6	106	0.00
67 C	Ethyl Benzene	1.857	1.821	1.9#	109	0.00
68 T	m/p-Xylenes	0.720	0.695	3.5	106	0.00
69 T	o-Xylene	0.675	0.660	2.2	107	0.00
70 T	Styrene	1.110	1.102	0.7	108	0.00
71 P	Bromoform	0.201	0.188	6.5	101	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	0.00
73 T	Isopropylbenzene	3.779	3.639	3.7	105	0.00
74 T	N-amyl acetate	0.655	0.654	0.2	107	0.00
75 P	1,1,2,2-Tetrachloroethane	0.567	0.538	5.1	103	0.00
76 T	1,2,3-Trichloropropane	0.386	0.409	-6.0	101	0.00
77 T	Bromobenzene	0.866	0.831	4.0	104	0.00
78 T	n-propylbenzene	4.400	4.232	3.8	105	0.00
79 T	2-Chlorotoluene	2.563	2.510	2.1	108	0.00
80 T	1,3,5-Trimethylbenzene	3.260	3.164	2.9	105	0.00
81 T	trans-1,4-Dichloro-2-butene	0.184	0.173	6.0	98	0.00
82 T	4-Chlorotoluene	2.729	2.658	2.6	107	0.00
83 T	tert-Butylbenzene	2.860	2.784	2.7	105	0.00
84 T	1,2,4-Trimethylbenzene	3.290	3.248	1.3	107	0.00
85 T	sec-Butylbenzene	3.942	3.772	4.3	104	0.00
86 T	p-Isopropyltoluene	3.585	3.439	4.1	104	0.00
87 T	1,3-Dichlorobenzene	1.720	1.663	3.3	107	0.00
88 T	1,4-Dichlorobenzene	1.720	1.663	3.3	107	0.00
89 T	n-Butylbenzene	3.287	3.141	4.4	104	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW122118\
Data File : VW007721.D
Acq On : 20 Dec 2018 02:30
Operator : SY/AP
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_W
LabSampleId :
VSTDCCC050

Quant Time: Dec 20 07:44:11 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W122118S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 20 06:30:23 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.629	0.601	4.5	104	0.00
91 T	1,2-Dichlorobenzene	1.540	1.490	3.2	106	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.116	0.104	10.3	94	0.00
93 T	1,2,4-Trichlorobenzene	1.069	1.031	3.6	105	0.00
94 T	Hexachlorobutadiene	0.571	0.528	7.5	102	0.00
95 T	Naphthalene	1.976	1.979	-0.2	104	0.00
96 T	1,2,3-Trichlorobenzene	0.915	0.879	3.9	103	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6