

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW122318\
 Data File : VW007805.D
 Acq On : 21 Dec 2018 19:49
 Operator : SY/AP
 Sample : J6392-05RE
 Misc : 6.55G/5ML/MSVOA W/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 KY001SB103-121118-(7-9)RE

Manual Integrations
 APPROVED

MMDadoda
 12/26/2018 10:59:12 AM

Quant Time: Dec 22 02:49:25 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W122118S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 20 06:30:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	44076	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	56576	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	28660	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	8111	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	21506	44.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.04%	
35) Dibromofluoromethane	7.88	113	19511	56.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.06%	
50) Toluene-d8	10.32	98	57065	43.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.72%	
62) 4-Bromofluorobenzene	12.62	95	12415	24.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	49.40%	
Target Compounds						
16) Acetone	4.12	43	7157	79.115	ug/l	99
19) Methyl tert-butyl Ether	5.41	73	1849m	2.962	ug/l	
68) m/p-Xylenes	11.84	106	930	2.252	ug/l	85
84) 1,2,4-Trimethylbenzene	13.26	105	1617	3.030	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW122318\
Data File : VW007805.D
Acq On : 21 Dec 2018 19:49
Operator : SY/AP
Sample : J6392-05RE
Misc : 6.55G/5ML/MSVOA W/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
KY001SB103-121118-(7-9)RE

Manual Integrations
APPROVED

MMDadoda
12/26/2018 10:59:12 AM

Quant Time: Dec 22 02:49:25 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W122118S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 20 06:30:23 2018
Response via : Initial Calibration

