

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW122318\
 Data File : VW007814.D
 Acq On : 21 Dec 2018 23:44
 Operator : SY/AP
 Sample : J6377-05RE
 Misc : 5.36G/5ML/MSVOA W/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 P001-SS001-0612-02RE

Quant Time: Dec 22 03:05:41 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W122118S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 20 06:30:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	16093	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	30733	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	24181	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	7305	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	7746	43.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.82%	
35) Dibromofluoromethane	7.88	113	5991	31.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	63.90%	
50) Toluene-d8	10.32	98	14992	21.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	42.42%	
62) 4-Bromofluorobenzene	12.62	95	4825	17.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	35.34%	
Target Compounds						
16) Acetone	4.12	43	4135	125.189	ug/l	94
20) Methylene Chloride	4.90	84	2250	8.371	ug/l	96
52) Toluene	10.38	92	1874	3.529	ug/l	97
68) m/p-Xylenes	11.84	106	972	2.790	ug/l	100
84) 1,2,4-Trimethylbenzene	13.25	105	868	1.806	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW122318\
Data File : VW007814.D
Acq On : 21 Dec 2018 23:44
Operator : SY/AP
Sample : J6377-05RE
Misc : 5.36G/5ML/MSVOA W/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
P001-SS001-0612-02RE

Quant Time: Dec 22 03:05:41 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W122118S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 20 06:30:23 2018
Response via : Initial Calibration

