

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW122318\
 Data File : VW007817.D
 Acq On : 22 Dec 2018 01:02
 Operator : SY/AP
 Sample : J6386-18RE
 Misc : 6.6G/5ML/MSVOA W/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 P001-SS007-1824-01RE

Quant Time: Dec 22 03:15:50 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W122118S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 20 06:30:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	40422	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	77723	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	52957	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	14484	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	29583	66.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	133.54%	
35) Dibromofluoromethane	7.88	113	22560	47.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.16%	
50) Toluene-d8	10.32	98	57619	32.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	64.46%	
62) 4-Bromofluorobenzene	12.62	95	16374	23.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	47.42%	
Target Compounds						
8) Diethyl Ether	3.68	74	320	1.787	ug/l	81
16) Acetone	4.12	43	13691	165.023	ug/l	97
25) 2-Butanone	7.17	43	2672	27.497	ug/l	98
52) Toluene	10.38	92	1822	1.357	ug/l	92
68) m/p-Xylenes	11.84	106	2041	2.675	ug/l	99
69) o-Xylene	12.16	106	1007	1.408	ug/l	98
84) 1,2,4-Trimethylbenzene	13.26	105	1126	1.182	ug/l	93
93) 1,2,4-Trichlorobenzene	15.14	180	815	2.632	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW122318\
Data File : VW007817.D
Acq On : 22 Dec 2018 01:02
Operator : SY/AP
Sample : J6386-18RE
Misc : 6.6G/5ML/MSVOA W/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
P001-SS007-1824-01RE

Quant Time: Dec 22 03:15:50 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W122118S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 20 06:30:23 2018
Response via : Initial Calibration

