

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW122318\
 Data File : VW007819.D
 Acq On : 22 Dec 2018 01:54
 Operator : SY/AP
 Sample : J6386-20
 Misc : 6.97G/5ML/MSVOA W/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 P001-SS008-0612-01

Quant Time: Mar 01 04:29:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W122118S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 20 06:30:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	51967	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	89493	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	56644	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	15252	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	29014	50.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.88%	
35) Dibromofluoromethane	7.88	113	24033	44.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.04%	
50) Toluene-d8	10.32	98	62748	30.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	60.98%	
62) 4-Bromofluorobenzene	12.62	95	17435	21.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	43.86%	
Target Compounds						
16) Acetone	4.12	43	19486	182.693	ug/l	99
64) Tetrachloroethene	10.86	164	815	2.032	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW122318\
Data File : VW007819.D
Acq On : 22 Dec 2018 01:54
Operator : SY/AP
Sample : J6386-20
Misc : 6.97G/5ML/MSVOA W/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
P001-SS008-0612-01

Quant Time: Mar 01 04:29:22 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W122118S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 20 06:30:23 2018
Response via : Initial Calibration

