

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW122318\
 Data File : VW007824.D
 Acq On : 22 Dec 2018 06:40
 Operator : SY/AP
 Sample : VW1223SBL02
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VW1223SBL02

Quant Time: Dec 24 00:01:46 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W122118S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 20 06:30:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	67810	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	115028	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	105573	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	52749	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	38689	52.06	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.12%	
35) Dibromofluoromethane	7.88	113	34490	49.15	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.30%	
50) Toluene-d8	10.32	98	130519	49.34	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.68%	
62) 4-Bromofluorobenzene	12.62	95	50554	49.47	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.94%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW122318\
Data File : VW007824.D
Acq On : 22 Dec 2018 06:40
Operator : SY/AP
Sample : VW1223SBL02
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VW1223SBL02

Quant Time: Dec 24 00:01:46 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W122118S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 20 06:30:23 2018
Response via : Initial Calibration

