


Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW122318\
 Data File : VW007822.D
 Acq On : 22 Dec 2018 04:30
 Operator : SY/AP
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W122118S.M
 Title : SW846 8260
 Last Update : Thu Dec 20 06:30:23 2018



AutoFind: Scans 1757, 1758, 1759; Background Corrected with Scan 1749

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.3	11774	PASS
75	95	30	60	50.3	34213	PASS
95	95	100	100	100.0	67984	PASS
96	95	5	9	6.8	4597	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	75.4	51293	PASS
175	174	5	9	8.4	4289	PASS
176	174	95	101	96.6	49538	PASS
177	176	5	9	6.6	3290	PASS