




Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW122322\
 Data File : VW025197.D
 Acq On : 23 Dec 2022 15:18
 Operator : SY/MD
 Sample : N6171-01RE
 Misc : 4.30g/10mL/MSVOA_W/SoIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 GBQC1

Manual IntegrationsAPPROVED

Quant Time: Dec 23 22:03:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120822SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Dec 23 21:59:05 2022
 Response via : Initial Calibration

Reviewed By :Krupa Patel 12/27/2022
 Supervised By :Mahesh Dadoda 12/27/2022



TIC: VW025197.D\data.ms

(22) 2-Butanone (T)

7.171min (+ 0.000) 3.12 ug/L

response 9215

Ion	Exp%	Act%
43.00	100.00	100.00
72.00	30.00	29.90
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW122322\
 Data File : VW025197.D
 Acq On : 23 Dec 2022 15:18
 Operator : SY/MD
 Sample : N6171-01RE
 Misc : 4.30g/10mL/MSVOA_W/SoIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 GBQC1

Manual IntegrationsAPPROVED

Quant Time: Dec 23 22:03:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120822SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Dec 23 21:59:05 2022
 Response via : Initial Calibration

Reviewed By :Krupa Patel 12/27/2022
 Supervised By :Mahesh Dadoda 12/27/2022



TIC: VW025197.D\data.ms

(22) 2-Butanone (T)

7.171min (+ 0.000) 6.33 ug/L m

response 18676

Ion	Exp%	Act%
43.00	100.00	100.00
72.00	30.00	14.75#
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\Vw122322\
 Data File : VW025197.D
 Acq On : 23 Dec 2022 15:18
 Operator : SY/MD
 Sample : N6171-01RE
 Misc : 4.30g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 GBQC1

Manual IntegrationsAPPROVED

Quant Time: Dec 23 22:03:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120822SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Dec 23 21:59:05 2022
 Response via : Initial Calibration

Reviewed By :Krupa Patel 12/27/2022
 Supervised By :Mahesh Dadoda 12/27/2022

Compound	R.T.	Q	Ion	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene	8.842	114		788516	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117		727046	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152		293799	25.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3	2.349	65		160942	17.706	ug/L	0.00
Spiked Amount 25.000	Range 30	- 150		Recovery =	70.840%		
7) Chloroethane-d5	2.885	69		122944	20.947	ug/L	0.00
Spiked Amount 25.000	Range 30	- 150		Recovery =	83.800%		
11) 1,1-Dichloroethene-d2	4.013	63		286444	14.384	ug/L	0.00
Spiked Amount 25.000	Range 45	- 110		Recovery =	57.520%		
21) 2-Butanone-d5	7.074	46		219304	96.102	ug/L	0.00
Spiked Amount 50.000	Range 20	- 135		Recovery =	192.200%#		
24) Chloroform-d	7.647	84		427667	21.012	ug/L	0.00
Spiked Amount 25.000	Range 40	- 150		Recovery =	84.040%		
26) 1,2-Dichloroethane-d4	8.305	65		252538	24.466	ug/L	0.00
Spiked Amount 25.000	Range 70	- 130		Recovery =	97.880%		
32) Benzene-d6	8.275	84		883742	20.727	ug/L	0.00
Spiked Amount 25.000	Range 20	- 135		Recovery =	82.920%		
36) 1,2-Dichloropropane-d6	9.274	67		311692	25.124	ug/L	0.00
Spiked Amount 25.000	Range 70	- 120		Recovery =	100.480%		
41) Toluene-d8	10.323	98		829585	21.459	ug/L	0.00
Spiked Amount 25.000	Range 30	- 130		Recovery =	85.840%		
43) trans-1,3-Dichloroprop...	10.579	79		128827	24.582	ug/L	0.00
Spiked Amount 25.000	Range 30	- 135		Recovery =	98.320%		
47) 2-Hexanone-d5	10.921	63		187178	102.833	ug/L	0.00
Spiked Amount 50.000	Range 20	- 135		Recovery =	205.660%#		
56) 1,1,2,2-Tetrachloroeth...	12.688	84		316810	34.581	ug/L	0.00
Spiked Amount 25.000	Range 45	- 120		Recovery =	138.320%#		
66) 1,2-Dichlorobenzene-d4	13.847	152		219922	22.084	ug/L	0.00
Spiked Amount 25.000	Range 75	- 120		Recovery =	88.320%		
Target Compounds							
13) Acetone	4.123	43		93320	39.783	ug/L	96
22) 2-Butanone	7.171	43		18676m	6.328	ug/L	
46) Tetrachloroethene	10.860	164		33963	3.955	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW122322\
Data File : VW025197.D
Acq On : 23 Dec 2022 15:18
Operator : SY/MD
Sample : N6171-01RE
Misc : 4.30g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
GBQC1

Manual IntegrationsAPPROVED

Quant Time: Dec 23 22:03:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120822SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Dec 23 21:59:05 2022
Response via : Initial Calibration

Reviewed By :Krupa Patel 12/27/2022
Supervised By :Mahesh Dadoda 12/27/2022

