

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW122324\
 Data File : VW031506.D
 Acq On : 23 Dec 2024 14:47
 Operator : SY/MD
 Sample : P5351-02
 Misc : 11.44g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 E2924

Quant Time: Dec 24 00:16:54 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121324SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Dec 24 00:12:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	1072559	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	851308	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	290903	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.357	65	270186	16.742	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	66.960%	
7) Chloroethane-d5	2.899	69	211422	17.710	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	70.840%	
11) 1,1-Dichloroethene-d2	4.021	65	115098	15.036	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery	=	60.160%	
21) 2-Butanone-d5	7.106	46	157535	51.835	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	103.660%	
24) Chloroform-d	7.648	84	442160	15.046	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	=	60.200%	
26) 1,2-Dichloroethane-d4	8.307	65	278697	18.884	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery	=	75.520%	
32) Benzene-d6	8.276	84	901390	16.520	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	=	66.080%	
36) 1,2-Dichloropropane-d6	9.270	67	275927	17.351	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	=	69.400%#	
41) Toluene-d8	10.319	98	750875	14.792	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	=	59.160%	
43) trans-1,3-Dichloroprop...	10.581	79	97762	13.333	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	=	53.320%	
47) 2-Hexanone-d5	10.922	63	115368	53.946	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	107.900%	
56) 1,1,2,2-Tetrachloroeth...	12.690	84	227090	22.744	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	=	90.960%	
66) 1,2-Dichlorobenzene-d4	13.849	152	154565	14.312	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	=	57.240%#	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW122324\
Data File : VW031506.D
Acq On : 23 Dec 2024 14:47
Operator : SY/MD
Sample : P5351-02
Misc : 11.44g/10mL/MSVOA_W/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
E2924

Quant Time: Dec 24 00:16:54 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121324SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Dec 24 00:12:57 2024
Response via : Initial Calibration

