

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW122418\
 Data File : VW007865.D
 Acq On : 23 Dec 2018 00:57
 Operator : SY/AP
 Sample : J6511-07
 Misc : 4.45G/5ML/MSVOA_W/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 SB-63(0-5)-GRAB

Quant Time: Dec 24 03:58:55 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W122118S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 20 06:30:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	5509	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	9354	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	2212	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	218	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	7727	127.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	255.94%	
35) Dibromofluoromethane	7.88	113	6485	113.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	227.28%	
50) Toluene-d8	10.32	98	5587	25.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	51.94%	
62) 4-Bromofluorobenzene	12.62	95	413	4.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	9.94%	
Target Compounds						
16) Acetone	4.12	43	1923	170.073	ug/l	99
20) Methylene Chloride	4.90	84	1178	16.623	ug/l #	86
52) Toluene	10.38	92	1830	11.321	ug/l	96
67) Ethyl Benzene	11.74	91	448	5.453	ug/l	92
68) m/p-Xylenes	11.84	106	1355	42.517	ug/l	100
69) o-Xylene	12.18	106	282	9.440	ug/l	58
84) 1,2,4-Trimethylbenzene	13.25	105	309	21.544	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_W\DATA\VW122418\
Data File : VW007865.D
Acq On : 23 Dec 2018 00:57
Operator : SY/AP
Sample : J6511-07
Misc : 4.45G/5ML/MSVOA_W/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
SB-63(0-5)-GRAB

Quant Time: Dec 24 03:58:55 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W122118S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 20 06:30:23 2018
Response via : Initial Calibration

