

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW122418\
 Data File : VW007867.D
 Acq On : 23 Dec 2018 01:49
 Operator : SY/AP
 Sample : J6511-09
 Misc : 5.92G/5ML/MSVOA_W/SOIL
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 SB-65(0-5)-GRAB

Quant Time: Dec 24 04:05:16 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W122118S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 20 06:30:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	74261	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	131491	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	116051	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	51824	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	43228	53.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.22%	
35) Dibromofluoromethane	7.88	113	37230	46.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.82%	
50) Toluene-d8	10.32	98	143654	47.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.00%	
62) 4-Bromofluorobenzene	12.62	95	53166	45.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.02%	
Target Compounds						
8) Diethyl Ether	3.67	74	649	1.972	ug/l	96
16) Acetone	4.12	43	980	6.430	ug/l	93
20) Methylene Chloride	4.91	84	6807	3.003	ug/l	92
52) Toluene	10.38	92	4574	2.013	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_W\DATA\VW122418\
Data File : VW007867.D
Acq On : 23 Dec 2018 01:49
Operator : SY/AP
Sample : J6511-09
Misc : 5.92G/5ML/MSVOA_W/SOIL
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
SB-65(0-5)-GRAB

Quant Time: Dec 24 04:05:16 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W122118S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 20 06:30:23 2018
Response via : Initial Calibration

