

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW122418\
 Data File : VW007874.D
 Acq On : 23 Dec 2018 06:34
 Operator : SY/AP
 Sample : J6446-02
 Misc : 5.35G/5ML/MSVOA_W/SOIL
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 P001-SS025-1218-02

Quant Time: Dec 24 15:15:11 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W122118S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 20 06:30:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	60233	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	92518	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	51069	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	12955	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	40231	60.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	121.88%	
35) Dibromofluoromethane	7.88	113	32069	56.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.64%	
50) Toluene-d8	10.32	98	87725	41.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.46%	
62) 4-Bromofluorobenzene	12.62	95	18097	22.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	44.04%	
Target Compounds						
16) Acetone	4.12	43	17225	139.332	ug/l	98
68) m/p-Xylenes	11.83	106	841	1.143	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_W\DATA\VW122418\
Data File : VW007874.D
Acq On : 23 Dec 2018 06:34
Operator : SY/AP
Sample : J6446-02
Misc : 5.35G/5ML/MSVOA_W/SOIL
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
P001-SS025-1218-02

Quant Time: Dec 24 15:15:11 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W122118S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 20 06:30:23 2018
Response via : Initial Calibration

