

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW122420\
 Data File : VW017735.D
 Acq On : 24 Dec 2020 14:50
 Operator : SY/VA
 Sample : L5206-01
 Misc : 4.41G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BG4L8

Manual Integrations
 APPROVED

MMDadoda
 12/26/2020 1:37:35 PM

Quant Time: Dec 25 08:01:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM122420S.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 25 07:46:43 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	428057	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	390525	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	176278	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	135371	20.44	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	81.76%
7) Chloroethane-d5	2.89	69	121224	22.18	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	88.72%
10) 1,1-Dichloroethene-d2	4.03	63	178533	15.47	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	61.88%
20) 2-Butanone-d5	7.08	46	55110	51.34	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	102.68%
24) Chloroform-d	7.65	84	254342	22.74	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	90.96%
26) 1,2-Dichloroethane-d4	8.30	65	142274	24.57	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	98.28%
29) Benzene-d6	8.27	84	504461	22.93	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	91.72%
33) 1,2-Dichloropropane-d6	9.27	67	147612	24.34	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	97.36%
37) Toluene-d8	10.32	98	461174	21.83	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	87.32%
38) trans-1,3-Dichloropropene-	10.58	79	58016	21.46	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	85.84%
39) 2-Hexanone-d5	10.93	63	39687	52.12	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	104.24%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	108411	24.08	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	96.32%
61) 1,2-Dichlorobenzene-d4	13.85	152	139669	22.75	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	91.00%

Target Compounds					Ovalue
13) Acetone	4.13	43	1611	1.738 ug/L	54
16) Methylene chloride	4.92	84	8966m	1.465 ug/L	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW122420\
Data File : VW017735.D
Acq On : 24 Dec 2020 14:50
Operator : SY/VA
Sample : L5206-01
Misc : 4.41G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BG4L8

Manual Integrations
APPROVED

MMDadoda
12/26/2020 1:37:35 PM

Quant Time: Dec 25 08:01:08 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM122420S.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 25 07:46:43 2020
Response via : Initial Calibration

