

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW122521\
 Data File : VW021647.D
 Acq On : 25 Dec 2021 15:44
 Operator : SY/VA
 Sample : M5174-09RE
 Misc : 4.79g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 GBC20RE

Quant Time: Dec 27 00:41:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120921SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Dec 24 01:43:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	54605	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.628	117	46007	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	17376	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	13745	13.821	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	55.280%
7) Chloroethane-d5	2.879	69	10996	16.725	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	66.920%
11) 1,1-Dichloroethene-d2	4.007	63	17995	13.572	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	54.280%
21) 2-Butanone-d5	7.086	46	10493	74.851	ug/L	-0.01
Spiked Amount	50.000	Range	20 - 135	Recovery	=	149.700%#
24) Chloroform-d	7.653	84	31205	20.506	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	82.040%
26) 1,2-Dichloroethane-d4	8.305	65	19088	23.582	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	94.320%
32) Benzene-d6	8.275	84	53278	21.127	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	84.520%
36) 1,2-Dichloropropane-d6	9.274	67	16543	24.460	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	97.840%
41) Toluene-d8	10.323	98	46828	18.174	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	72.680%
43) trans-1,3-Dichloroprop...	10.579	79	6117	19.643	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	78.560%
47) 2-Hexanone-d5	10.927	63	8228	82.273	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	164.540%#
56) 1,1,2,2-Tetrachloroeth...	12.688	84	17535	30.160	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	120.640%#
66) 1,2-Dichlorobenzene-d4	13.853	152	12161	18.059	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	72.240%#
Target Compounds						
					Qvalue	
13) Acetone	4.141	43	2020	16.074	ug/L	86
33) Benzene	8.329	78	2033	0.811	ug/L	100
34) Trichloroethene	9.092	95	3745	5.323	ug/L	93
42) Toluene	10.390	91	2848	0.972	ug/L	98
46) Tetrachloroethene	10.866	164	14790	22.130	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW122521\
Data File : VW021647.D
Acq On : 25 Dec 2021 15:44
Operator : SY/VA
Sample : M5174-09RE
Misc : 4.79g/10.0mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
GBC20RE

Quant Time: Dec 27 00:41:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120921SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Dec 24 01:43:02 2021
Response via : Initial Calibration

