

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW122521\
 Data File : VW021657.D
 Acq On : 25 Dec 2021 19:46
 Operator : SY/VA
 Sample : VIBLK
 Misc : 5.00g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK569

Quant Time: Dec 27 00:42:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120921SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Dec 24 01:43:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	107415	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.634	117	106720	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	65615	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	37535	19.187	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	76.760%
7) Chloroethane-d5	2.879	69	23332	18.041	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	72.160%
11) 1,1-Dichloroethene-d2	4.019	63	47489	18.208	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	72.840%
21) 2-Butanone-d5	7.092	46	18093	65.611	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	131.220%
24) Chloroform-d	7.653	84	67805	22.651	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	90.600%
26) 1,2-Dichloroethane-d4	8.305	65	39086	24.547	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	98.200%
32) Benzene-d6	8.274	84	132899	22.719	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	90.880%
36) 1,2-Dichloropropane-d6	9.274	67	36531	23.286	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	93.160%
41) Toluene-d8	10.323	98	132802	22.219	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	88.880%
43) trans-1,3-Dichloroprop...	10.579	79	15859	21.955	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	87.800%
47) 2-Hexanone-d5	10.926	63	15646	67.444	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	134.880%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	35122	26.043	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	104.160%
66) 1,2-Dichlorobenzene-d4	13.853	152	55704	21.905	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	87.640%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW122521\
Data File : VW021657.D
Acq On : 25 Dec 2021 19:46
Operator : SY/VA
Sample : VIBLK
Misc : 5.00g/10.0mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK569

Quant Time: Dec 27 00:42:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120921SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Dec 24 01:43:02 2021
Response via : Initial Calibration

