


Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW122618\
Data File : VW007903.D
Acq On : 24 Dec 2018 22:20
Operator : SY/AP
Sample : BFB
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W122618S.M
Title : SW846 8260
Last Update : Tue Dec 25 04:35:47 2018



AutoFind: Scans 1757, 1758, 1759; Background Corrected with Scan 1749

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.4	12888	PASS
75	95	30	60	51.7	36320	PASS
95	95	100	100	100.0	70200	PASS
96	95	5	9	6.8	4751	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	73.5	51578	PASS
175	174	5	9	8.2	4212	PASS
176	174	95	101	95.8	49394	PASS
177	176	5	9	6.7	3312	PASS