

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW122718\
 Data File : VW007928.D
 Acq On : 26 Dec 2018 17:32
 Operator : SY/AP
 Sample : J6471-15
 Misc : 5.20G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 F9H02

Quant Time: Dec 27 01:50:59 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM122618S.M
 Quant Title : VOC Analysis
 QLast Update : Thu Dec 27 00:43:27 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	73591	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	60436	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	25614	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	17955	24.76	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	99.04%
7) Chloroethane-d5	2.90	69	12164	24.89	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	99.56%
10) 1,1-Dichloroethene-d2	4.01	63	35520	18.09	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	72.36%
20) 2-Butanone-d5	7.08	46	13187	42.64	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	85.28%
24) Chloroform-d	7.65	84	44207	22.58	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	90.32%
26) 1,2-Dichloroethane-d4	8.31	65	29244	24.91	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	99.64%
29) Benzene-d6	8.27	84	86464	25.51	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	102.04%
33) 1,2-Dichloropropane-d6	9.27	67	24301	26.36	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	105.44%
37) Toluene-d8	10.32	98	78724	23.84	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	95.36%
38) trans-1,3-Dichloropropene-	10.58	79	12165	22.98	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	91.92%
39) 2-Hexanone-d5	10.93	63	11432	48.48	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	96.96%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	21243	23.94	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	95.76%
61) 1,2-Dichlorobenzene-d4	13.86	152	22951	22.18	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	88.72%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW122718\
Data File : VW007928.D
Acq On : 26 Dec 2018 17:32
Operator : SY/AP
Sample : J6471-15
Misc : 5.20G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
F9H02

Quant Time: Dec 27 01:50:59 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM122618S.M
Quant Title : VOC Analysis
QLast Update : Thu Dec 27 00:43:27 2018
Response via : Initial Calibration

