

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW122718\
 Data File : VW007983.D
 Acq On : 27 Dec 2018 21:20
 Operator : SY/AP
 Sample : J6540-08
 Misc : 4.21G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 GAEN9

Quant Time: Dec 28 06:38:26 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM122618S.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 28 05:56:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	96196	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	91186	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	45119	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	22721	23.97	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	95.88%
7) Chloroethane-d5	2.90	69	17292	27.07	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	108.28%
10) 1,1-Dichloroethene-d2	4.01	63	46290	18.04	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	72.16%
20) 2-Butanone-d5	7.08	46	23250	57.51	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	115.02%
24) Chloroform-d	7.65	84	68472	26.75	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	107.00%
26) 1,2-Dichloroethane-d4	8.31	65	45650	29.75	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	119.00%
29) Benzene-d6	8.27	84	128050	25.04	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	100.16%
33) 1,2-Dichloropropane-d6	9.27	67	36337	26.12	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	104.48%
37) Toluene-d8	10.32	98	121001	24.29	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	97.16%
38) trans-1,3-Dichloropropene-	10.58	79	20634	25.84	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	103.36%
39) 2-Hexanone-d5	10.93	63	20705	58.19	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	116.38%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	38125	28.48	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	113.92%
61) 1,2-Dichlorobenzene-d4	13.86	152	47980	26.33	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	105.32%

Target Compounds					Ovalue
13) Acetone	4.12	43	4430	11.215 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW122718\
Data File : VW007983.D
Acq On : 27 Dec 2018 21:20
Operator : SY/AP
Sample : J6540-08
Misc : 4.21G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
GAEN9

Quant Time: Dec 28 06:38:26 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM122618S.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 28 05:56:01 2018
Response via : Initial Calibration

