

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW122719\
 Data File : VW014486.D
 Acq On : 28 Dec 2019 03:02
 Operator : SY/VA
 Sample : K6397-02
 Misc : 5.03G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK01

Quant Time: Dec 28 04:37:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM122719S.M
 Quant Title : VOC Analysis
 QLast Update : Sat Dec 28 03:24:24 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	1214277	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	1083688	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	505483	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	415156	24.25	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	97.00%
7) Chloroethane-d5	2.89	69	307008	24.34	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	97.36%
10) 1,1-Dichloroethene-d2	4.02	63	564354	18.02	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	72.08%
20) 2-Butanone-d5	7.07	46	262948	64.55	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	129.10%
24) Chloroform-d	7.65	84	719823	25.77	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	103.08%
26) 1,2-Dichloroethane-d4	8.30	65	399473	27.55	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	110.20%
29) Benzene-d6	8.27	84	1622578	28.16	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	112.64%
33) 1,2-Dichloropropane-d6	9.27	67	490222	28.29	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	113.16%
37) Toluene-d8	10.32	98	1481083	28.35	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	113.40%
38) trans-1,3-Dichloropropene-	10.57	79	197607	27.61	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	110.44%
39) 2-Hexanone-d5	10.92	63	211166	71.27	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	142.54%#
48) 1,1,2,2-Tetrachloroethane-	12.69	84	395283	30.38	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	121.52%#
61) 1,2-Dichlorobenzene-d4	13.85	152	501843	30.13	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	120.52%#


Target Compounds					Ovalue
13) Acetone	4.12	43	22090	3.917 ug/L	69


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW122719\
Data File : VW014486.D
Acq On : 28 Dec 2019 03:02
Operator : SY/VA
Sample : K6397-02
Misc : 5.03G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK01

Quant Time: Dec 28 04:37:09 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM122719S.M
Quant Title : VOC Analysis
QLast Update : Sat Dec 28 03:24:24 2019
Response via : Initial Calibration





#13

Acetone

Concen: 3.917 ug/L

RT: 4.12 min Scan# 363

Delta R.T. -0.00 min

Lab File: VW014486.D

Acq: 28 Dec 2019 03:02

Instrument :
MSVOA_W
ClientSampleId :
VHBLK01

Tot Ion: 43 Resp: 22090

Ion Ratio Lower Upper

43 100

58 14.9 0.0 64.0

