


Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW122818\
 Data File : VW007993.D
 Acq On : 28 Dec 2018 07:21
 Operator : SY/AP
 Sample : BFB82
 Misc : 5.00G/10ML/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BFB82

Integration File: LSCINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM122618S.M
 Title : VOC Analysis
 Last Update : Fri Dec 28 22:34:20 2018



AutoFind: Scans 1757, 1758, 1759; Background Corrected with Scan 1750

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.4	6040	PASS
75	95	30	80	53.3	17496	PASS
95	95	100	100	100.0	32832	PASS
96	95	5	9	6.8	2239	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	72.1	23680	PASS
175	174	5	9	8.6	2044	PASS
176	174	95	101	96.1	22765	PASS
177	176	5	9	6.8	1546	PASS

m/z Abundance

Instrument :
MSVOA_W
ClientSampleId :
BFB82

m/z	Abundance
36.00	287.0
37.10	1617.0
38.10	1387.0
39.10	533.0
44.10	215.0
45.10	411.0
47.10	201.0
47.90	224.0
49.10	1410.0
50.00	6020.0
51.10	1784.0
52.00	71.0
54.90	79.0
56.00	475.0
57.00	899.0
59.90	277.0
61.10	1723.0
62.00	1653.0
63.00	1252.0
64.10	164.0
64.90	146.0
66.90	86.0
68.10	2892.0
69.10	3224.0
70.00	224.0
72.10	159.0
73.10	1985.0
74.00	5564.0
75.10	16120.0
76.00	1704.0
77.10	256.0
78.20	99.0
79.00	1027.0
80.00	283.0
81.00	1004.0
81.80	224.0
87.00	845.0
88.00	1129.0
91.00	170.0
92.10	983.0
93.10	1591.0
94.10	3338.0
95.10	30800.0
96.10	1922.0
97.20	61.0
103.00	95.0
103.90	185.0
105.00	51.0
105.90	156.0
109.90	68.0
115.10	59.0
115.90	176.0
116.90	247.0
117.90	197.0
119.10	169.0
125.30	73.0
127.90	116.0
129.80	131.0
133.00	451.0
133.90	90.0
134.20	103.0
135.00	53.0
136.90	55.0
140.90	365.0
142.90	336.0
146.10	73.0
147.00	87.0
147.90	114.0
155.10	64.0
163.00	64.0
165.00	66.0
172.20	156.0
174.00	19776.0
175.10	1964.0
176.00	19168.0
177.00	1347.0
178.00	103.0
178.90	139.0
191.00	395.0
192.20	95.0
193.00	516.0
194.10	143.0
194.70	81.0
205.10	92.0
207.10	271.0
249.00	343.0
250.10	57.0
251.10	84.0
265.10	359.0
266.00	106.0
267.10	66.0
281.10	3471.0

Instrument :
MSVOA_W
ClientSampleId :
BFB82

282.00	1025.0
283.20	574.0
284.20	133.0

Instrument :
MSVOA_W
ClientSampleId :
BFB82

m/z	Abundance
36.10	314.0
37.10	1788.0
38.10	1535.0
39.10	599.0
44.10	238.0
45.00	383.0
47.20	350.0
47.90	272.0
49.00	1729.0
50.10	7045.0
51.10	2151.0
52.00	134.0
55.00	140.0
56.10	601.0
57.00	1162.0
60.10	459.0
61.10	1660.0
62.10	1845.0
63.00	1249.0
64.10	150.0
65.10	124.0
67.00	115.0
68.00	3677.0
69.10	3814.0
70.10	309.0
72.10	180.0
73.10	2103.0
74.00	6351.0
75.10	19800.0
76.10	1697.0
76.90	202.0
77.90	122.0
78.90	1251.0
79.90	344.0
80.90	1249.0
81.90	218.0
87.00	1424.0
88.10	1190.0
91.00	172.0
92.00	1097.0
93.10	1719.0
94.00	4789.0
95.10	37264.0
96.10	2384.0
97.10	98.0
103.10	105.0
104.00	178.0
104.90	70.0
106.00	213.0
109.90	58.0
112.90	62.0
115.10	89.0
115.90	171.0
117.00	264.0
117.80	163.0
118.90	244.0
125.40	55.0
127.80	161.0
128.90	89.0
129.90	140.0
133.10	367.0
133.90	83.0
135.00	87.0
136.70	73.0
140.90	408.0
142.90	472.0
146.00	95.0
147.70	104.0
154.90	93.0
160.90	77.0
164.90	85.0
172.20	179.0
174.00	26496.0
175.00	2151.0
176.00	25304.0
177.00	1609.0
177.90	92.0
178.90	121.0
191.20	269.0
192.20	100.0
193.00	464.0
194.00	89.0
194.70	89.0
205.10	82.0
207.00	304.0
249.10	278.0
251.10	111.0
265.00	281.0
266.00	68.0
267.10	76.0
281.10	3084.0
282.10	894.0

Instrument :
MSVOA_W
ClientSampleId :
BFB82

283.00488.0

284.10112.0

Instrument :
MSVOA_W
ClientSampleId :
BFB82

m/z	Abundance
36.10	291.0
37.10	1395.0
38.10	1211.0
39.00	459.0
44.00	228.0
45.00	195.0
47.10	190.0
48.10	199.0
49.10	1415.0
50.10	5056.0
51.10	1763.0
51.90	81.0
55.10	106.0
56.10	521.0
57.10	776.0
60.00	360.0
61.10	1482.0
62.10	1386.0
63.10	1193.0
64.00	192.0
66.80	61.0
68.00	2937.0
69.10	3212.0
70.10	270.0
72.10	176.0
73.10	1680.0
74.10	5324.0
75.10	16568.0
76.10	1423.0
77.10	260.0
78.00	125.0
78.90	1073.0
80.10	270.0
81.00	1037.0
81.80	188.0
87.00	962.0
87.80	916.0
88.10	816.0
91.10	80.0
92.00	847.0
93.10	1468.0
94.10	3715.0
95.10	30432.0
96.10	2411.0
103.90	106.0
104.90	56.0
106.10	114.0
115.00	56.0
116.10	133.0
116.90	210.0
117.80	182.0
119.10	229.0
125.40	70.0
127.70	129.0
128.80	78.0
129.90	115.0
133.20	188.0
135.00	66.0
141.00	442.0
142.90	399.0
144.90	53.0
145.90	65.0
147.90	117.0
155.10	56.0
160.90	64.0
172.10	131.0
174.00	24768.0
175.00	2017.0
176.00	23824.0
177.00	1683.0
178.10	72.0
178.90	61.0
190.90	215.0
192.10	105.0
192.90	269.0
207.20	148.0
248.90	120.0
251.10	75.0
265.00	110.0
265.30	112.0
281.10	1890.0
282.20	544.0
283.10	344.0
284.10	72.0

Instrument :
MSVOA_W
ClientSampleId :
BFB82