

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW122821\
 Data File : VW021684.D
 Acq On : 28 Dec 2021 21:59
 Operator : SY/VA
 Sample : M5171-17
 Misc : 5.01g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK001

Quant Time: Dec 29 00:27:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM122821SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Dec 29 00:22:49 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	88399	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.634	117	86112	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	48390	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.343	65	47882	27.472	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 109.880%	
7) Chloroethane-d5	2.879	69	28384	24.179	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 96.720%	
11) 1,1-Dichloroethene-d2	4.013	63	44315	17.374	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	= 69.480%	
21) 2-Butanone-d5	7.092	46	12974	50.879	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 101.760%	
24) Chloroform-d	7.647	84	65849	23.847	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	= 95.400%	
26) 1,2-Dichloroethane-d4	8.305	65	36146	24.790	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	= 99.160%	
32) Benzene-d6	8.275	84	126537	24.747	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	= 99.000%	
36) 1,2-Dichloropropane-d6	9.275	67	34344	24.760	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	= 99.040%	
41) Toluene-d8	10.323	98	122251	23.997	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	= 96.000%	
43) trans-1,3-Dichloroprop...	10.579	79	16810	24.906	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	= 99.640%	
47) 2-Hexanone-d5	10.927	63	11323	53.826	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 107.660%	
56) 1,1,2,2-Tetrachloroeth...	12.689	84	30290	26.488	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	= 105.960%	
66) 1,2-Dichlorobenzene-d4	13.847	152	47157	25.484	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	= 101.920%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW122821\
Data File : VW021684.D
Acq On : 28 Dec 2021 21:59
Operator : SY/VA
Sample : M5171-17
Misc : 5.01g/10.0mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK001

Quant Time: Dec 29 00:27:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM122821SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Dec 29 00:22:49 2021
Response via : Initial Calibration

