

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW123021\
 Data File : VW021706.D
 Acq On : 30 Dec 2021 14:32
 Operator : SY/VA
 Sample : M5194-21
 Misc : 5.02g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK001

Quant Time: Dec 31 00:53:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM122821SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Dec 31 00:51:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	79910	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.634	117	78934	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	46271	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	41849	26.561	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 106.240%	
7) Chloroethane-d5	2.885	69	28343	26.709	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 106.840%	
11) 1,1-Dichloroethene-d2	4.019	63	40682	17.644	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	= 70.560%	
21) 2-Butanone-d5	7.080	46	12987	56.341	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 112.680%	
24) Chloroform-d	7.653	84	60017	24.043	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	= 96.160%	
26) 1,2-Dichloroethane-d4	8.305	65	33181	25.174	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	= 100.680%	
32) Benzene-d6	8.275	84	113720	24.263	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	= 97.040%	
36) 1,2-Dichloropropane-d6	9.275	67	31875	25.070	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	= 100.280%	
41) Toluene-d8	10.323	98	112983	24.194	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	= 96.760%	
43) trans-1,3-Dichloroprop...	10.579	79	14881	24.053	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	= 96.200%	
47) 2-Hexanone-d5	10.927	63	10298	53.405	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 106.820%	
56) 1,1,2,2-Tetrachloroeth...	12.689	84	27820	26.540	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	= 106.160%	
66) 1,2-Dichlorobenzene-d4	13.853	152	41959	23.713	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	= 94.840%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW123021\
Data File : VW021706.D
Acq On : 30 Dec 2021 14:32
Operator : SY/VA
Sample : M5194-21
Misc : 5.02g/10.0mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK001

Quant Time: Dec 31 00:53:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM122821SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Dec 31 00:51:50 2021
Response via : Initial Calibration

