

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW123021\
 Data File : VW021710.D
 Acq On : 30 Dec 2021 16:54
 Operator : SY/VA
 Sample : M5214-10
 Misc : 5.00g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK001

Quant Time: Dec 31 00:54:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM122821SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Dec 31 00:51:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.847	114	72077	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.634	117	74202	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	43409	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	41355	29.100	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 116.400%	
7) Chloroethane-d5	2.885	69	29197	30.504	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 122.000%	
11) 1,1-Dichloroethene-d2	4.019	63	39395	18.943	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	= 75.760%	
21) 2-Butanone-d5	7.073	46	12212	58.736	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 117.480%	
24) Chloroform-d	7.646	84	60216	26.745	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	= 106.960%	
26) 1,2-Dichloroethane-d4	8.305	65	32588	27.411	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	= 109.640%	
32) Benzene-d6	8.274	84	113989	25.872	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	= 103.480%	
36) 1,2-Dichloropropane-d6	9.274	67	31175	26.083	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	= 104.320%	
41) Toluene-d8	10.323	98	110876	25.257	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	= 101.040%	
43) trans-1,3-Dichloroprop...	10.579	79	15003	25.796	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	= 103.200%	
47) 2-Hexanone-d5	10.920	63	9587	52.889	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 105.780%	
56) 1,1,2,2-Tetrachloroeth...	12.688	84	27778	28.190	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	= 112.760%	
66) 1,2-Dichlorobenzene-d4	13.853	152	42912	25.851	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	= 103.400%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW123021\
Data File : VW021710.D
Acq On : 30 Dec 2021 16:54
Operator : SY/VA
Sample : M5214-10
Misc : 5.00g/10.0mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK001

Quant Time: Dec 31 00:54:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM122821SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Dec 31 00:51:50 2021
Response via : Initial Calibration

