

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW123021\
 Data File : VW021709.D
 Acq On : 30 Dec 2021 16:30
 Operator : SY/VA
 Sample : M5125-04
 Misc : 5.02g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK001

Quant Time: Dec 31 00:54:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM122821SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Dec 31 00:51:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	80792	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.633	117	80496	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	47087	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	40537	25.448	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 101.800%	
7) Chloroethane-d5	2.879	69	27680	25.799	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 103.200%	
11) 1,1-Dichloroethene-d2	4.013	63	39565	16.972	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	= 67.880%	
21) 2-Butanone-d5	7.073	46	13334	57.215	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 114.420%	
24) Chloroform-d	7.646	84	61632	24.421	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	= 97.680%	
26) 1,2-Dichloroethane-d4	8.305	65	33476	25.120	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	= 100.480%	
32) Benzene-d6	8.274	84	115121	24.086	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	= 96.360%	
36) 1,2-Dichloropropane-d6	9.274	67	31790	24.518	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	= 98.080%	
41) Toluene-d8	10.323	98	113131	23.756	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	= 95.040%	
43) trans-1,3-Dichloroprop...	10.579	79	15266	24.196	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	= 96.800%	
47) 2-Hexanone-d5	10.920	63	10363	52.700	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 105.400%	
56) 1,1,2,2-Tetrachloroeth...	12.694	84	28602	26.757	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	= 107.040%	
66) 1,2-Dichlorobenzene-d4	13.853	152	42609	23.663	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	= 94.640%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW123021\
Data File : VW021709.D
Acq On : 30 Dec 2021 16:30
Operator : SY/VA
Sample : M5125-04
Misc : 5.02g/10.0mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK001

Quant Time: Dec 31 00:54:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM122821SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Dec 31 00:51:50 2021
Response via : Initial Calibration

