

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW010522\
 Data File : VW021741.D
 Acq On : 05 Jan 2022 19:21
 Operator : SY/VA
 Sample : M5342-05
 Misc : 3.30g/5.0mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 SB-8-(0-2)

Quant Time: Jan 06 05:54:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W010522S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 05 14:56:42 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.945	168	68966	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.842	114	123312	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.628	117	41051	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.554	152	4285	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.311	65	48629	80.004	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 160.000%	
35) Dibromofluoromethane	7.885	113	54597	71.138	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 142.280%	
50) Toluene-d8	10.323	98	92698	33.372	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 66.740%	
62) 4-Bromofluorobenzene	12.615	95	6828	6.622	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 13.240%#	
Target Compounds						
						Qvalue
20) Methylene Chloride	4.916	84	17945	17.036	ug/l	90
44) Trichloroethene	9.098	130	2020	2.141	ug/l	86
64) Tetrachloroethene	10.860	164	2933	10.356	ug/l	90
68) m/p-Xylenes	11.829	106	1229	2.138	ug/l	92
78) n-propylbenzene	12.810	91	477	1.377	ug/l #	68
80) 1,3,5-Trimethylbenzene	12.938	105	435	1.741	ug/l	87
84) 1,2,4-Trimethylbenzene	13.255	105	1022	4.179	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW010522\
Data File : VW021741.D
Acq On : 05 Jan 2022 19:21
Operator : SY/VA
Sample : M5342-05
Misc : 3.30g/5.0mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
SB-8-(0-2)

Quant Time: Jan 06 05:54:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W010522S.M
Quant Title : SW846 8260
QLast Update : Wed Jan 05 14:56:42 2022
Response via : Initial Calibration

