

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW010522\
 Data File : VW021745.D
 Acq On : 05 Jan 2022 20:57
 Operator : SY/VA
 Sample : M5342-09
 Misc : 6.25g/5.0mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 SB-10-(0-2)

Quant Time: Jan 06 05:55:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W010522S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 05 14:56:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.945	168	161628	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.841	114	254214	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.628	117	228017	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.554	152	103081	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.311	65	70266	49.326	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	98.660%
35) Dibromofluoromethane	7.884	113	73468	46.434	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	92.860%
50) Toluene-d8	10.323	98	257648	44.993	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	89.980%
62) 4-Bromofluorobenzene	12.615	95	86642	40.761	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	81.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW010522\
Data File : VW021745.D
Acq On : 05 Jan 2022 20:57
Operator : SY/VA
Sample : M5342-09
Misc : 6.25g/5.0mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
SB-10-(0-2)

Quant Time: Jan 06 05:55:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W010522S.M
Quant Title : SW846 8260
QLast Update : Wed Jan 05 14:56:42 2022
Response via : Initial Calibration

