

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW011521\
 Data File : VW017914.D
 Acq On : 15 Jan 2021 21:41
 Operator : SY/VA
 Sample : VIBLK27
 Misc : 5.00G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK27

Manual Integrations
 APPROVED

apatel
 1/18/2021 8:51:02 AM

Quant Time: Jan 18 04:38:50 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM122420S.M
 Quant Title : VOC Analysis
 QLast Update : Mon Jan 18 02:07:49 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	383417	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	353372	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	171690	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	116102	19.57	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	78.28%
7) Chloroethane-d5	2.89	69	106758	21.80	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	87.20%
10) 1,1-Dichloroethene-d2	4.02	63	169043	16.36	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	65.44%
20) 2-Butanone-d5	7.09	46	55939	58.18	ug/L	0.01
Spiked Amount	50.000	Range	20 - 135	Recovery	=	116.36%
24) Chloroform-d	7.65	84	245229	24.48	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	97.92%
26) 1,2-Dichloroethane-d4	8.31	65	137197	26.46	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	105.84%
29) Benzene-d6	8.27	84	479384	24.08	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	96.32%
33) 1,2-Dichloropropane-d6	9.27	67	138785	25.29	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	101.16%
37) Toluene-d8	10.32	98	436217	22.82	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	91.28%
38) trans-1,3-Dichloropropene-	10.58	79	52500	21.47	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	85.88%
39) 2-Hexanone-d5	10.93	63	39720	57.65	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	115.30%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	107521	26.40	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	105.60%
61) 1,2-Dichlorobenzene-d4	13.85	152	149602	25.02	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	100.08%

Target Compounds				Ovalue
16) Methylene chloride	4.92	84	19894m	3.628 ug/L

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW011521\
Data File : VW017914.D
Acq On : 15 Jan 2021 21:41
Operator : SY/VA
Sample : VIBLK27
Misc : 5.00G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1


Instrument :
MSVOA_W
ClientSampleId :
VIBLK27

Manual Integrations
APPROVED

apatel
1/18/2021 8:51:02 AM

Quant Time: Jan 18 04:38:50 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM122420S.M
Quant Title : VOC Analysis
QLast Update : Mon Jan 18 02:07:49 2021
Response via : Initial Calibration





#16

Methylene chloride

Concen: 3.628 ug/L m

RT: 4.92 min Scan# 495

Delta R.T. 0.01 min

Lab File: VW017914.D

Acq: 15 Jan 2021 21:41

Tot Ion:	84	Resp:	19894
Ion Ratio	Lower	Upper	
84	100		
49	121.2	81.2	150.8
86	64.1	44.5	82.7

Instrument :
MSVOA_W
ClientSampleId :
VIBLK27

Manual Integrations
APPROVED

apatel
1/18/2021 8:51:02 AM

