

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW022819\
 Data File : VW008908.D
 Acq On : 01 Mar 2019 02:16
 Operator : SY/VA
 Sample : K1643-01
 Misc : 5.10G/5ML/MSVOA W/SOIL
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 INTERIOR-WALL

Quant Time: Mar 01 07:07:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W022119S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 22 00:38:32 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	295970	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	460504	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	407302	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	193263	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	164793	59.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.72%	
35) Dibromofluoromethane	7.88	113	139609	50.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.98%	
50) Toluene-d8	10.32	98	548830	51.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.84%	
62) 4-Bromofluorobenzene	12.62	95	187734	44.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.84%	
Target Compounds						
16) Acetone	4.14	43	5966	9.272	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW022819\
Data File : VW008908.D
Acq On : 01 Mar 2019 02:16
Operator : SY/VA
Sample : K1643-01
Misc : 5.10G/5ML/MSVOA W/SOIL
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
INTERIOR-WALL

Quant Time: Mar 01 07:07:10 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W022119S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 22 00:38:32 2019
Response via : Initial Calibration

