


Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW022819\
Data File : VW008902.D
Acq On : 28 Feb 2019 22:41
Operator : SY/VA
Sample : BFB
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W022119S.M
Title : SW846 8260
Last Update : Fri Feb 22 00:38:32 2019



AutoFind: Scans 1757, 1758, 1759; Background Corrected with Scan 1748

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.8	23280	PASS
75	95	30	60	51.8	60853	PASS
95	95	100	100	100.0	117538	PASS
96	95	5	9	6.6	7725	PASS
173	174	0.00	2	0.8	827	PASS
174	95	50	100	84.8	99682	PASS
175	174	5	9	8.2	8163	PASS
176	174	95	101	97.0	96642	PASS
177	176	5	9	7.0	6743	PASS