

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW032521\
 Data File : VW018450.D
 Acq On : 25 Mar 2021 11:40
 Operator : SY/VA
 Sample : VW0325SBL01
 Misc : 5.00G/10ML/MSVOA_W/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK12

Quant Time: Mar 26 01:30:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SOM2WLM031921S.M
 Quant Title : VOC Analysis
 QLast Update : Fri Mar 26 01:29:58 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	446337	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	401403	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.560	152	186502	25.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	141463	22.78	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	91.12%
7) Chloroethane-d5	2.892	69	132582	24.39	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	97.56%
10) 1,1-Dichloroethene-d2	4.019	63	169814	15.83	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	63.32%
20) 2-Butanone-d5	7.074	46	66747	48.27	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	96.54%
24) Chloroform-d	7.647	84	264490	23.27	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	93.08%
26) 1,2-Dichloroethane-d4	8.305	65	164397	24.77	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	99.08%
29) Benzene-d6	8.275	84	515477	24.16	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	96.64%
33) 1,2-Dichloropropane-d6	9.275	67	151546	24.62	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	98.48%
37) Toluene-d8	10.323	98	488879	23.79	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	95.16%
38) trans-1,3-Dichloroprop...	10.579	79	72382	22.80	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	91.20%
39) 2-Hexanone-d5	10.921	63	56159	46.62	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	93.24%
48) 1,1,2,2-Tetrachloroeth...	12.689	84	132637	24.09	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	96.36%
61) 1,2-Dichlorobenzene-d4	13.853	152	162777	24.55	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	98.20%
Target Compounds						
16) Methylene chloride	4.916	84	4776	0.65	ug/L	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW032521\
Data File : VW018450.D
Acq On : 25 Mar 2021 11:40
Operator : SY/VA
Sample : VW0325SBL01
Misc : 5.00G/10ML/MSVOA_W/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK12

Quant Time: Mar 26 01:30:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SOM2WLM031921S.M
Quant Title : VOC Analysis
QLast Update : Fri Mar 26 01:29:58 2021
Response via : Initial Calibration

