

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW072720\
 Data File : VW015970.D
 Acq On : 27 Jul 2020 16:36
 Operator : SY/VA
 Sample : L3458-06
 Misc : 5.09G/10ML/MSVOA W/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Jul 28 02:47:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM072020S.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jul 28 01:36:11 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	394511	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	363790	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	168684	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	90422	16.24	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	64.96%
7) Chloroethane-d5	2.89	69	71600	15.84	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	63.36%
10) 1,1-Dichloroethene-d2	4.01	63	135993	12.08	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	48.32%
20) 2-Butanone-d5	7.09	46	113725	77.77	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	155.54%#
24) Chloroform-d	7.65	84	185090	16.70	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	66.80%
26) 1,2-Dichloroethane-d4	8.31	65	118592	19.50	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	78.00%
29) Benzene-d6	8.28	84	370780	16.74	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	66.96%
33) 1,2-Dichloropropane-d6	9.28	67	119296	17.97	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	71.88%
37) Toluene-d8	10.33	98	330239	16.46	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	65.84%
38) trans-1,3-Dichloropropene-	10.58	79	53846	19.03	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	76.12%
39) 2-Hexanone-d5	10.93	63	88066	79.67	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	159.34%#
48) 1,1,2,2-Tetrachloroethane-	12.69	84	145226	28.46	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	113.84%
61) 1,2-Dichlorobenzene-d4	13.85	152	123439	18.77	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	75.08%

Target Compounds

					Ovalue
13) Acetone	4.13	43	19578	15.943	ug/L 97
15) Methyl Acetate	4.67	43	5949m	2.569	ug/L
16) Methylene chloride	4.92	84	14377	2.224	ug/L 90
21) 2-Butanone	7.18	43	3232	2.128	ug/L 99
22) cis-1,2-Dichloroethene	7.18	96	48650	8.683	ug/L 94
35) Trichloroethene	9.10	95	136335	23.326	ug/L 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW072720\
Data File : VW015970.D
Acq On : 27 Jul 2020 16:36
Operator : SY/VA
Sample : L3458-06
Misc : 5.09G/10ML/MSVOA W/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :

Quant Time: Jul 28 02:47:45 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM072020S.M
Quant Title : VOC Analysis
QLast Update : Tue Jul 28 01:36:11 2020
Response via : Initial Calibration

